




Advanced Hydrogen Sensing



## OPERATING MANUAL

# HY-ALERTA® 5000 Series Hydrogen Sensor

27215 Turnberry Lane, Unit A

Valencia, CA. 91355, USA

Tel: (661) 775-9575

E-mail: [hello@h2scan.com](mailto:hello@h2scan.com)

Website: <http://www.h2scan.com>

## **IMPORTANT NOTICES**



Read and understand this operating manual before installing or using the unit. If this equipment is used in a manner not specified by H2scan, the protection provided by this equipment may be impaired.

**LIMITATION OF LIABILITY** - seller shall under no circumstances be liable for any incidental, consequential, special, punitive, or other damages, including, but not limited to, loss of business or profit, promotional or manufacturing expenses, injury to reputation, or loss of customer, based on any alleged negligence, breach of warranty, strict liability, breach of contract, or any other legal theory arising out of the use, misuse, purchase, sale or possession of its goods or its performance of this contract to the extent that such liability extends seller's obligations beyond the price paid by buyer to seller for the item on which such claim is based. Seller advises buyer to perform acceptable tests on all hardware prior to deployment and to perform maintenance as described in the seller's instruction guide. Under no circumstances shall the equipment provided hereunder be used in a manner where it is the sole protective system for facilities, equipment, and personnel safety; the equipment is intended for use in conjunction with other appropriate protective systems.

## **LIMITED WARRANTY**

**H2scan Limited Warranty:** Each hydrogen instrument ("Product") will conform, as to all substantial operational features, to the Product specifications set forth in this Manual and will be free of defects which substantially affect such Product's performance for twelve (12) months from the ship date for such Product.

**Must Provide Notice of Defect:** If you believe a Product that you believe is defective, you must notify H2scan in writing, within ten (10) days of receipt of such Product, of your claim regarding any such defect.

**Return Product to H2scan for Repair, Replacement or Credit:** If the Product is found defective by H2scan, H2scan's sole obligation under this warranty is to either (i) repair the Product, (ii) replace the Product, or (iii) issue a credit for the purchase price for such Product, the remedy to be determined by H2scan on a case-by-case basis.

**Voided Warranty:** H2scan's 12 Month Limited Warranty is void for any of the following:

- The unit is opened, and the manufacturing seal is broken.
- Unauthorized repair work performed at the customer's location or conducted by anyone other than H2scan's factory trained technicians.
- Equipment or parts that have been tampered with, misused, neglected, mishandled, improperly adjusted, or modified in any way without the written consent of H2scan.
- Equipment or parts that have been damaged due to shipping, misuse, accidents, mishandling, neglect, or problems with electrical power sources.
- Repair work performed during the warranty period does not prolong the warranty period past the original period.
- System operation in incorrect or inappropriate environments.
- Usage that is not in accordance with system guidelines or an operator's failure to follow manual instructions.

**LIMITATION OF WARRANTY:** THE ABOVE IS A LIMITED WARRANTY AS IT IS THE ONLY WARRANTY MADE BY H2SCAN. H2SCAN MAKES NO OTHER WARRANTY EXPRESS OR IMPLIED AND EXPRESSLY EXCLUDES ALL WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. YOUR SOLE REMEDY HEREUNDER IS REPAIR OR REPLACEMENT OF THE PRODUCT OR A CREDIT FOR THE PURCHASE PRICE FOR SUCH PRODUCT, THE PARTICULAR REMEDY TO BE DETERMINED BY H2SCAN ON A CASE-BY-CASE BASIS. H2SCAN SHALL HAVE NO LIABILITY WITH RESPECT TO ITS OBLIGATIONS UNDER THIS AGREEMENT FOR CONSEQUENTIAL, EXEMPLARY, OR INCIDENTAL DAMAGES EVEN IF IT HAS BEEN ADVISED OF THE POSSIBILITY OF SUCH DAMAGES. THE STATED EXPRESS WARRANTY IS IN LIEU OF ALL LIABILITIES OR OBLIGATIONS OF H2SCAN FOR DAMAGES ARISING OUT OF OR IN CONNECTION WITH THE DELIVERY, USE OR PERFORMANCE OF THE PRODUCTS.

## Contents

<b>1</b>	<b>Introduction</b>	<b>5</b>
<b>2</b>	<b>Features</b>	<b>6</b>
2.1	SENSOR .....	6
2.2	MECHANICAL .....	6
2.3	PHYSICAL BARRIER .....	7
2.4	ELECTRICAL FEATURES .....	7
<b>3</b>	<b>Specifications</b>	<b>8</b>
3.1	PERFORMANCE SPECIFICATIONS .....	8
3.2	CERTIFICATIONS .....	9
3.3	STANDARDS .....	9
<b>4</b>	<b>Electrical Interface</b>	<b>10</b>
4.1	CONNECTOR .....	10
4.2	WIRING DIAGRAM .....	10
4.3	CABLE .....	10
4.4	POWER SUPPLY .....	11
4.5	RS485 .....	11
<b>5</b>	<b>Installation</b>	<b>12</b>
5.1	HANDLING PRECAUTIONS .....	12
5.2	MOUNTING .....	12
5.3	ELECTRICAL CONNECTION .....	12
5.4	OPTIONAL ACCESSORIES .....	12
5.4.1	Grounding Lug .....	12
5.4.2	Mounting Bracket .....	12
5.4.3	Conduit Adapter .....	13
5.4.4	5902 HY-ALERTA Analog Output Module .....	13
5.4.5	Custom Sensor Cable Lengths .....	13
<b>6</b>	<b>Operation</b>	<b>13</b>
6.1	STARTUP .....	13
6.2	MONITORING .....	13
6.3	ERROR/EXCEPTION HANDLING .....	13
<b>7</b>	<b>2-Point Field Calibration</b>	<b>14</b>
7.1	CALIBRATION GASES .....	14
7.2	BACKGROUND GASES .....	14
7.3	FIELD CALIBRATION PROCEDURE .....	14
7.4	ABORT FIELD CALIBRATION .....	16
7.5	CLEAR FIELD CALIBRATION .....	16
7.6	CALIBRATION STATUS AND ERRORS .....	16
<b>8</b>	<b>Modbus</b>	<b>18</b>
8.1	COMMUNICATION SETTINGS .....	18
8.2	PROTOCOL .....	18
8.2.1	Exception Response .....	19
8.2.2	Modbus Command Register Definitions .....	20
8.2.3	Hydrogen Measurement .....	24
8.2.4	Temperature Measurement .....	24
8.2.5	ASCII Strings .....	24
8.2.6	Date Register Format .....	25

8.2.7	Status and Error Information .....	25
8.2.8	Set Unit ID.....	26
8.2.9	Stop Bit Selection .....	26
8.2.10	Baud Rate .....	27
8.2.11	Real time Clock .....	27
8.2.12	User Information.....	27
<b>9</b>	<b>Firmware Upgrade .....</b>	<b>28</b>
<b>10</b>	<b>Principles of Operation .....</b>	<b>29</b>
10.1	EFFECT OF PRESSURE.....	29
10.2	PRESSURE NORMALIZED SPECIFICATIONS .....	30
10.3	SENSOR BEHAVIOR.....	30
10.4	NON-REDUCING AND NON-OXIDIZING ATMOSPHERES .....	31
10.5	DIURNAL EFFECT .....	31
10.6	RESPONSE TIME .....	31
10.7	SENSOR PERFORMANCE EVALUATION (RESPONSE TIME) .....	31
10.8	SENSOR PERFORMANCE EVALUATION (ACCURACY).....	32
10.9	GENERAL TROUBLESHOOTING .....	32

## 1 Introduction

The HY-ALERTA 5000 Series Hydrogen Sensors measure hydrogen concentration in mixed gases using H2scan's patented solid-state, non-consumable, hydrogen-specific sensing element. H2scan's patented technology provides a reliable, real-time, direct, hydrogen-specific measurement without cross sensitivities to common industrial gases such as hydrocarbons, carbon dioxide, and inert gases. The robust design remains accurate over time without periodic calibration thereby reducing services cost.

This product series is ideal for real-time inline process applications to improve efficiency/yields and to provide safe and effective operations in petrochemical processing and hydrogen economy applications such as electrolyzers, fuel cells, and hydrogen distribution, blending and storage.

Based on H2scan's 5<sup>th</sup> generation electronics platform, the HY-ALERTA 5000 Series Hydrogen Sensors report calibrated hydrogen readings through a digital interface. H2scan's solid-state hydrogen sensor and patented algorithms are combined in the HY-ALERTA 5000 to provide continuous, maintenance free, dependable, and accurate hydrogen measurement.

The HY-ALERTA 5000 Series Hydrogen Sensors have a single electrical connection to provide DC power and 2-wire, RS-485 communications using Modbus RTU protocol.

This document has been updated for use with firmware revision **3:5:A**.

## 2 Features

### 2.1 Sensor

- The sensor measures and reports hydrogen as its partial pressure and therefore can measure hydrogen as a constituent of various gas mixtures.
- The sensor does not require maintenance.
- The sensing element has a long life and is not consumed or degraded during operation.
- There are no moving parts to wear out.
- No background gas, carrier gas, or any other consumables required for operation.
- The sensors support a field calibration function.

### 2.2 Mechanical

The HY-ALERTA 5000 Series Hydrogen Sensors have a rugged, waterproof mechanical assembly design for industrial applications. A  $\frac{3}{4}$ "-14 NPT fitting is provided for either the attachment of the included protective sensor cap for standard area monitoring applications or can be used to directly install the sensor into a cabinet or other gas compartment. The electronics packaging of the HY-ALERTA 5000 is IP68 and saltwater corrosion rated for marine applications\*. (Exceeds C5M requirements.) Dimensions of various configurations are shown below.




Figure 1: Hydrogen Sensor with Protective Cap Dimensions (Standard HY-ALERTA 5000 Series Hydrogen Sensor Configuration)






Figure 2: Hydrogen Sensor Dimensions (Sensor Only)

\* See Section 3 Specifications for information relating to limitations of these ratings.



**Figure 3: Hydrogen Sensor in Mounting Bracket Dimensions**



**Figure 4: Hydrogen Sensor with Protective Sensor Cap, Grounding Clip, Mounting Bracket, and Conduit Adapter Dimensions**

## 2.3 Physical Barrier

A physical barrier exists between the hydrogen sensing element and the HY-ALERTA 5000 Series Hydrogen Sensor electronics. This barrier includes a glass-to-metal feedthrough for electrical connections and gaskets to complete the hydrogen gas-tight seal.

## 2.4 Electrical Features

The HY-ALERTA 5000 Series Hydrogen Sensor has a single 4-pin M12 connector for power and communications.

- DC power input of 9 to 48 volts, 10 watts (commonly available 24 VDC or 48 VDC power supplies are recommended)
- 2-wire RS485 for Modbus RTU communications

### 3 Specifications

**Table 1: Absolute Maximum Operating Conditions**

Parameter	Value			Units
	Minimum	Nominal	Maximum	
<b>Environment – Gas Stream<sup>†</sup></b>				
Operating Temperature	-40		60	°C
Pressure <sup>‡</sup> (absolute)	0.1	1	10	atm abs.
Operating Humidity	< 95% RH (non-condensing)			
<b>Environment – Ambient</b>				
Operating Temperature	-40		70	°C
Storage Temperature	-40		105	°C
Ingress Protection <sup>§</sup>	IP68; 25 feet water for 14 days (IEC 60529)			
Humidity	0 to 100% RH, condensing <sup>§</sup>			
Corrosion Resistance <sup>§</sup>	Class C5M Marine rated; salt-water condensing (IEC 60068-2-11 & DIN EN ISO 12944)			
<b>Mechanical</b>				
Vibration	3-axis Sinusoidal, Wideband and Random (IEC 60068-2-6 table C.2, IEC 60068-2-64 paragraph A.2, category no. 2, IEC 61373: 2010 Cat 1B section 9)			
Shock	30g, shock duration 18ms (IEC 60068-2-27)			
Weight	0.85 lbs. (387 grams)			
<b>Electrical</b>				
Voltage Input	8.1	24	52.8	VDC
Power Consumption			10	W

#### 3.1 Performance Specifications

Sensor performance specifications are absolute, and assume an ambient temperature of 25°C, 1 atmosphere absolute pressure, and are in addition to any errors in the calibration gases used. Accuracy and repeatability are defined as  $\pm$  the values listed. For other operating pressures, see Section 10.2 *Pressure Normalized Specifications* as it will affect the specifications listed in the tables below.

While the sensors can operate in gas containing no hydrogen, measurement performance specifications are valid only when the sensor is exposed to gas containing hydrogen above the Lower Detection Limit shown below.

<sup>†</sup> Specific product models may have different limits.

<sup>‡</sup> Operating at pressures above or below 1 atm abs. will affect the performance. See 10.1 Effect of Pressure.

<sup>§</sup> In addition to other material limitations described in this document, the sensing elements cannot be exposed to condensing water or corrosive materials. To achieve IP68 and saltwater corrosion rating for marine applications for the sensing elements, the mechanical connection of the unit must be installed into an environment that is barriered from environment containing the condensing water and/or corrosive materials with equivalent or better ingress protection and corrosion resistance as stated.

**Table 2: Model-Dependent Performance Specifications**

<b>MODEL</b>	<b>H<sub>2</sub> Range</b>		<b>Lower Detection Limit (LDL)</b>	<b>Response Time**</b>	<b>Accuracy (Absolute Error)</b>	<b>Repeatability (Absolute Error)</b>
	<b>Low</b>	<b>High</b>				
<b>HY-ALERTA 5020</b>	0%	5%	0.4% H <sub>2</sub>	< 60 sec	0.3% H <sub>2</sub>	0.3% H <sub>2</sub>
<b>HY-ALERTA 5021</b>	0%	5%	0.4% H <sub>2</sub>	< 5 min	0.3% H <sub>2</sub>	0.3% H <sub>2</sub>

### 3.2 Certifications



IP68

### 3.3 Standards

- IEC 60068-2-2 & EN 50155 Section 13.4.4
- IEC 60068-2-11 & DIN EN ISO 12944
- IEC 60529
- IEC 60068-2-6 table C.2
- IEC 60068-2-64 paragraph A.2, category no. 2
- IEC 60068-2-27
- FCC Part 15
- EN 55011 Class A Group 1
- IEC 61000-4-2, 61000-4-3, 61000-4-4, 61000-4-6, and 61000-4-8
- ANSI/UL/IEC/EN 61010-1
- IEC 61326-1

\*\* Response time is defined as the maximum amount of time for the unit to report a hydrogen concentration greater than 2% when the gas *at the sensor* is transitioned from 0% to 3% H<sub>2</sub>/Air at 1 atmosphere absolute of pressure. For a more complete discussion on response time, please refer to section 10.6 *Response Time*.

## 4 Electrical Interface

All electrical connections to the HY-ALERTA 5000 Series Hydrogen Sensors are supplied through a single 4-pin M12 connector.

### 4.1 Connector

The key (notch) location and pin numbers are shown in **Table 3**.


**Table 3: HY-ALERTA 5000 Series Hydrogen Sensor Pin Out**

Pin	Signal Name	Wire Color
1	DC power	Brown
2	DC ground	White
3	RS485 Data+	Blue
4	RS485 Data-	Black

**Note:** This view is looking at the connector on the HY-ALERTA 5000 Series Hydrogen Sensor.

### 4.2 Wiring Diagram

A wiring diagram is shown in **Figure 5**.



**Figure 5: Connection to Modbus RTU Master**

### 4.3 Cable

A suitably rated cable assembly is required for most field applications. A 5m cable is included as standard, all other lengths are custom orderable options.

Cable recommendations are:

- 4-pin M12 Female molded connector
- IP68 rated connector
- 4/18 AWG wire
- Shielded cable is recommended (**grounded at one end only**)

**Note:** H2scan is not responsible for issues arising from the use of cabling and connectors that are not suitable for the environment.

#### 4.4 Power Supply

The HY-ALERTA 5000 Series Hydrogen Sensors incorporate an internal isolated voltage regulator for operation in harsh electrical conditions. The DC ground line is electrically isolated from the metal case. The HY-ALERTA 5000 Series Hydrogen Sensor enclosure is connected to earth ground through the mechanical connection.

Use an industrial grade, fixed-output power supply that meets the following specifications:

- DC voltage output
- Output Voltage: 12, 15, 24, 30, 36 or 48 VDC nominal (8.55 to 50.4 VDC)
- Output power: 10 watts minimum

**Note:** 24 VDC or 48 VDC Power Supply is recommended.

A 24 VDC Power Supply is available as an orderable accessory from H2scan.

#### 4.5 RS485

The RS485 input is galvanically isolated inside the HY-ALERTA 5000 Series Hydrogen Sensor to improve noise immunity in harsh electrical environments.

The HY-ALERTA 5000 Series Hydrogen Sensors do not include bias resistors for the Data lines which may need to be added at the SCADA end of the cable. A  $120\ \Omega$  termination resistor, between Data+ and Data-, is installed in the HY-ALERTA 5000 Series Hydrogen Sensor.

The following communication settings are used for the RS485, 2-wire, half-duplex connection:

Baud Rate:	9,600
	14,400
	19,200 (Factory Default)
	38,400
	57,600
	115,200

Data bits: 8

Stop bits: 1 or 2 (default 2)

Parity: None

Flow Control: None

## 5 Installation

### 5.1 Handling Precautions

The following precautions must be followed to ensure the sensor assembly is not damaged during handling:

- Ensure that nothing comes in contact with the sensor-end of the device.
- The protective cap should remain in place until the time of installation.
- Place a wrench on the metal housing closest to the threaded end when tightening. The HY-ALERTA 5000 Series Hydrogen Sensors are designed to handle torque throughout the whole sensor assembly enclosure.

### 5.2 Mounting

HY-ALERTA 5000 Series Hydrogen Sensors must be mounted such that the sensor tube is not pointed upward as this may cause accumulation of condensed fluids in the sensor cavity.

The sensor can be installed into a  $\frac{3}{4}$ "-14 NPT fitting for monitoring cabinets or other gas compartments. Use standard torque for  $\frac{3}{4}$ " NPT fitting which is approximately 50 ft-lbs. Do not overtighten.

**WARNING:** Mounting the sensor pointing upwards can cause the accumulation of condensed liquids in the sensor cavity and can cause issues with sensor operation and reporting and can void the limited warranty.

### 5.3 Electrical Connection

Before energizing the equipment:

- Install the cable between the HY-ALERTA 5000 Series Hydrogen Sensor and the power/communication equipment
- Communication ground must be connected to DC ground
- Connect the cable to the power/communication equipment
- Connect the cable to the HY-ALERTA 5000 Series Hydrogen Sensor

**Note:** The connector to the HY-ALERTA 5000 Series Hydrogen Sensor must be fully tightened to ensure the IP68 rating.

Turn on power and check Modbus communications. The default communication settings are:

- RS485, half-duplex, 19200 baud, 8 data bits, 2 stop bits, no parity
- Default Modbus ID is 1

### 5.4 Optional Accessories

H2scan offers accessories for the HY-ALERTA 5000 Series Hydrogen Sensors to assist with integration and installation of the hydrogen sensor. For more information about these accessories, check the [h2scan.com](http://h2scan.com) website and email [help@h2scan.com](mailto:help@h2scan.com).

#### 5.4.1 Grounding Lug

The ground lug provides a secure earth ground connection to HY-ALERTA 5000 Series Hydrogen Sensor.

#### 5.4.2 Mounting Bracket

The mounting bracket provides a method of attaching the HY-ALERTA 5000 Series Hydrogen Sensor to a surface.

#### 5.4.3 Conduit Adapter

The conduit adapter can be fixed to the end of the hydrogen sensor with the M12 connector, providing a feedthrough for the sensor cable through a 1/2" NPT conduit connection.

#### 5.4.4 5902 HY-ALERTA Analog Output Module

The analog output module converts the digital Modbus signal from the HY-ALERTA 5000 Series Hydrogen Sensor into a 4-20mA analog signal.

#### 5.4.5 Custom Sensor Cable Lengths

A standard 5 m cable is included with every HY-ALERTA 5000 Series Hydrogen Sensor. If longer lengths are necessary, custom cables to 30 m can be ordered from H2scan.

### 6 Operation

**Note:** If operating at a pressure other than 1 atm absolute, the hydrogen reading will need to be compensated for pressure by performing a 2-point field calibration or by dividing the hydrogen reading by the known pressure in atm absolute.

#### 6.1 Startup

Power on the analyzer by connecting the sensor cable and turning on the power supply. The sensor will execute a startup sequence and will begin reporting hydrogen after 1 minute. During this period, the analyzer will:

- Perform a power on system self-test
- Restore configuration settings from non-volatile memory.

The Modbus status register **111** bit **15** will indicate **Ready** when the first valid hydrogen measurement is available. After the startup sequence completes, measured and calculated values will be available in the Modbus registers.

If the analyzer has been powered off for more than one month, the real time clock may need to be set. See 8.2.11 Real time Clock.

If an error is reported, turn off power to the sensor; double check the electrical connections and power supply voltage before restoring power again. If the error condition persists, contact H2scan customer service for assistance at [help@h2scan.com](mailto:help@h2scan.com).

#### 6.2 Monitoring

During normal operation, the HY-ALERTA 5000 Series Hydrogen Sensor measurements should be polled through the RS-485 interface periodically for a measurement reading. The time between readings can be from 1 second to several hours or days depending on user requirements. Each reading should include the following Modbus holding registers.

- **Status Register (111 bits 15 & 12)** – Bit 15 indicates that the hydrogen measurement is available. Bit 12 indicates that there is an error.
- **Error Status Registers (112,113)** – Indicates which error is detected. (These registers are active when register 111 bit 12 is high)
- **Hydrogen Registers (0,1)** – Provides the Hydrogen ppm values. Programming note: The high word (0) must be read to enable the low word (1) value to be available.

#### 6.3 Error/Exception Handling

The HY-ALERTA 5000 Series Hydrogen Sensors are designed for continuous operation and will automatically recover from intermittent problems due to insufficient power, excessive electrical noise, or excessive internal PCB temperature.

If the sensing element is damaged and unable to operate, the HY-ALERTA 5000 Series Hydrogen Sensor will shut down the measurement system and continue responding to Modbus for error reporting. This error will be reported via register **111** bit **12** and then details specified in register **112,113**. This type of error typically indicates a hardware fault that can only be repaired at H2scan. Power cycle the unit to attempt recovery. If the error condition persists, contact H2scan for repair at [help@h2scan.com](mailto:help@h2scan.com).

## 2-Point Field Calibration

The HY-ALERTA 5000 Series Hydrogen Sensors support a 2-point field calibration. By applying two known gas concentrations, the sensor system will calculate and apply a gain and offset to its measurements to match its hydrogen reading with the known gas concentrations. This can be used to normalize the hydrogen reading when operating the sensor at a pressure other than 1 atmosphere absolute (see section 10.1 Effect of Pressure).

### 6.4 Calibration Gases

**WARNING:** DO NOT use gases with a hydrogen concentration exceeding 5% H<sub>2</sub> at 1.0 atm

The recommended gases for the HY-ALERTA 5000 Series Hydrogen Sensors are 1% H<sub>2</sub> in air and 2% H<sub>2</sub> in air.

The calibration gases **must** contain hydrogen above the lower detection limit of the sensor (0.4% H<sub>2</sub> at 1 atmosphere absolute).

Two gases are required. The given accuracy specification of the analyzer does not include the error of the calibration gas certification and is only valid with an exposure time of at least 60 minutes per gas.

Gases are applied to the analyzer through user's plumbing. Make sure the system allows for calibration gases to be applied at the same pressure as the analyzer sees during normal operation.

### 6.5 Background Gases

Calibration of the HY-ALERTA 5000 Series Hydrogen Sensors should be done with gas standards that are made with hydrogen diluted in one of the following gases:

- Air (recommended)
- N<sub>2</sub>
- Hydrocarbons (Alkanes / Alkenes / Alkynes)
- CO<sub>2</sub>
- He, Ar, etc. (Noble)

Contact H2scan for inquiries regarding potential background gases not listed above.

### 6.6 Field Calibration Procedure

**NOTE:** The exposure time per gas is a minimum of 60 minutes per exposure. Exposures shorter than 60 minutes can generate erroneous calibration data.

**Two** calibration hydrogen gas standards are required: one high and one low.

Do NOT use gases exceeding 5% H<sub>2</sub> concentration

Do NOT use 100% air, 100% N<sub>2</sub>, or any gas with an H<sub>2</sub> concentration below the LDL of the product model

**NOTE:** The calibration can be aborted at any time (See: 7.4 Abort Field Calibration).

**Table 4: 2-Point Field Calibration Registers**

Register	Parameter	Function	Data Type	Data Range	Access
126	Calibration Gas 1, ppm H <sub>2</sub> *	High word	32-bit binary number	0 to 10,000,000	R/W
127		Low word			
128	Calibration Complete/Date	High byte: Month Low byte: Day	32-bit binary value		R/W
129		Year			
130	Calibration Gas 2, ppm H <sub>2</sub> *	High word	32-bit binary number	0 to 10,000,000	R/W
131		Low word			
132	Field Cal start command	Gas exposure duration (minutes)	16-bit binary number	60 to 1440	W
133	Field Cal abort command	Abort Field Cal or clear data	8-bit binary number	1: abort 2: clear	W
134	Field Cal get status command	High byte: Error code Low byte: Status	16-bit binary number	Table 5 and Table 6	R
135	Gas start command	Gas 1 or 2 started	8-bit binary number	1: gas 1 2: gas 2	W

\*Note: To convert % to ppm, multiply % reading by 10,000.

The status and error codes should be read from register **134** after each write to the field calibration registers. Once the calibration is initiated, the status code should read '1' (in-progress) and the error code should read '0' (no errors detected) after register write until the procedure is completed. Once completed, the status code should read '0' (success) and the error code should read '0' (no errors detected). Refer to 7.6 Calibration Status and Errors if an unexpected code appears.

**Note:** For best results, remove any oxide layer on the sensing elements which may have formed by applying the high span calibration gas to the sensor for 60 minutes before initiating the calibration procedure below.

1. Initiate the calibration by writing the duration per gas exposure in minutes to register **132** (the minimum time is 60 minutes per gas with longer being better up to 1440 minutes).
2. Apply the first gas to the unit at 1 slpm.
3. Adjust the pressure to match the process flow to be measured.
4. Indicate that the first gas is flowing, by writing '1' to register **135**.
5. Wait the exposure time.
6. Write the first gas concentration in **ppm** to registers **126, 127**.

**Tip:** To convert % to ppm, multiply % value by 10,000.

7. Apply the second gas at 1 slpm.
8. Adjust the pressure to match the process flow to be measured.
9. Indicate that the second gas is flowing by writing '2' to register **135**.
10. Wait the exposure time.
11. Write the second gas concentration in **ppm** to registers **130, 131**.
12. Finish the field calibration with the current date, by writing to registers **128, 129**.

**Note:** If the entered calibration date does not match the current date in the unit, the field calibration will not complete. The unit's current date can be found in registers 175, 176. See Table 13 for formatting details.

13. The unit will automatically reboot.

## 6.7 Abort Field Calibration

If a field calibration is in progress but needs to be aborted, this can be performed by either power cycling the unit or writing '1' to register 133.

Register	Parameter	Function	Data Type	Data Range	Access
133	Field Cal abort command	Abort Field Cal or clear data	8-bit binary number	1: abort 2: clear	W

## 6.8 Clear Field Calibration

Clearing a field calibration removes the active field calibration. To clear the field calibration, write '2' to register 133.

**Note:** An analyzer operating without an active field calibration will measure hydrogen using the factory calibration. Unless a special factory calibration was requested, this is performed at an operating pressure of 1 atmosphere absolute. If the analyzer is operated at any other pressure the hydrogen measurement will need to be compensated. See section 10.1 Effect of Pressure.

Register	Parameter	Function	Data Type	Data Range	Access
133	Field Cal abort command	Abort Field Cal or clear data	8-bit binary number	1: abort 2: clear	W

## 6.9 Calibration Status and Errors

To see calibration status and errors, read from Register 134.

Register	Parameter	Function	Data Type	Data Range	Access
134	Field Cal get status command	High byte: Error code Low byte: Status	16-bit binary number	Table 5 and Table 6	R

**Table 5: Field Calibration Status**

<b>Code</b>	<b>Description</b>	<b>Notes</b>
0	Success	The calibration completed successfully.
1	In-progress	The calibration process is on-going.
2	Fail	The calibration stopped due to errors.
3	Abort	The calibration was cancelled by the user.
4	Clear	The calibration corrections have been removed.
5	Bad Input	The most recent command was not executed. See the error code for more details.

**Table 6: Field Calibration Errors**

<b>Code</b>	<b>Description</b>	<b>Notes</b>
0	No errors	No errors detected
2	Field calibration already in-progress	The command to start a field calibration was given but a field calibration is already in progress. The field calibration in progress must first be aborted.
3	Bad gas exposure time	The entered time is outside the exposure time limits. Ensure the time was entered in minutes and within the limits of the product (minimum of 60 minutes with more being better).
4	Timer error	There was an internal problem with the unit; power cycle and try again. Contact H2scan if this problem persists.
5	N/A	This code is used for diagnostic purposes only.
6	Gain error	The calculated calibration gain is outside the limits. Check that the gas was flowing, and the gas concentrations were entered in ppm and match the respective gases.
7	Bad gas concentration	The entered gas concentration is outside the limits for the product. Ensure the gas concentration is entered in ppm, and within the range of the product model.
8	Bad gas number	The command to start Gas 2 was received before Gas 1 completed.
9	Data not ready	A command to finish the calibration was received before both calibration points were completed.
10	Calculation failed	A gain and offset could not be generated based on the collected data.
11	Calibration date mismatch	The entered date must match the date stored in the unit to the day. Note that the unit's date stored in registers 175, 176 is in a different format than that expected in registers 128, 129. See Table 13 for details.
12	Bad result	The calculated gain and/or offset are outside the limits. Check that the gas was flowing, and the gas concentrations were entered in ppm and match the respective gases.
13	Bad state	A step was skipped in the calibration process.
14	Exposure not done	The gas exposure was not completed. Retry the command once the full exposure time has elapsed. If an incorrect gas exposure time was entered, abort the field calibration, and start over.

## 7 Modbus

The HY-ALERTA 5000 Series Sensors use Modbus RTU to communicate with external equipment. Modbus RTU is a popular industrial interface supported by many products.

### 7.1 Communication Settings

Modbus protocol communicates over RS485 and supports RTU packets. The HY-ALERTA 5000 Series Hydrogen Sensors' default Modbus ID is 1. The Modbus ID can be changed by writing to holding register **150**.

### 7.2 Protocol

The following tables comprise the list of Modbus packets, values, registers, and register definitions.

The maximum time for the sensor to respond to a Modbus command is 10 seconds. Therefore, the master's timeout should be set to 10,000 milliseconds or greater.

**Table 7: Modbus Read Request Packet**

Byte	Modbus Parameter	Range	Meaning
1	Slave address	1 – 247	Unit ID Address
2	Function Code	03	Read Holding Register
3	Starting Address Hi	0x00 – 0xFF	Holding Register Hi Byte
4	Starting Address Lo	0x00 – 0xFF	Holding Register Lo Byte
5	Number of registers Hi	0	Limited by Modbus spec V1.1b
6	Number of Registers Lo	1 – 125	Number of 16-bit registers Lo Byte
7	CRC Lo	0x00 – 0xFF	CRC Low Byte
8	CRC Hi	0x00 – 0xFF	CRC High Byte

**Table 8: Modbus Read Response Packet**

Byte	Modbus Parameter	Range	Meaning
1	Slave address	1 – 247	Unit ID Address
2	Function Code	03	Returning Holding Registers
3	Byte Count	7 – 255	Number of data bytes returned = N
4	1st Data Value Hi	0x00 – 0xFF	
5	1st Data Value Lo	0x00 – 0xFF	
6	2nd Data Value Hi	0x00 – 0xFF	
7	2nd Data Value Lo	0x00 – 0xFF	
...	...		
...	...		
2N+4	CRC Lo	0x00 – 0xFF	CRC Low Byte
2N+5	CRC Hi	0x00 – 0xFF	CRC High Byte

N is the number of bytes returned based on the number of registers requested. If N registers are requested, then 2N+5 bytes are returned.

**Table 9: Modbus Write Request Packet**

Byte	Modbus Parameter	Range	Meaning
1	Slave address	1 – 247	Unit ID Address
2	Function Code	06	Write Holding Registers
3	Register Address Hi	0x00 – 0xFF	Unit Register Address Hi byte
4	Register Address Lo	0x00 – 0xFF	Unit Register Address Lo byte
5	Data Value Hi Byte	0x00 – 0xFF	
6	Data Value Lo Byte	0x00 – 0xFF	
7	CRC Lo	0x00 – 0xFF	CRC Low Byte
8	CRC Hi	0x00 – 0xFF	CRC High Byte

**Table 10: Modbus Write Response Packet**

Byte	Modbus Parameter	Range	Meaning
1	Slave address	1 – 247	Unit ID Address
2	Function Code	06	
3	Register Address Hi Byte	0x00 – 0xFF	Unit Register Address Hi byte
4	Register Address Lo Byte	0x00 – 0xFF	Unit Register Address Lo byte
5	Data Value Hi Byte	0x00 – 0xFF	
6	Data Value Lo Byte	0x00 – 0xFF	
7	CRC Lo	0x00 – 0xFF	CRC Low Byte
8	CRC Hi	0x00 – 0xFF	CRC High Byte

### 7.2.1 Exception Response

In a normal communications query and response due to a communication error, the master device sends a query to the slave device. Upon receiving the query, the slave processes the request and returns a response to the master device. An abnormal communication between the two devices produces one of four possible events:

1. If the slave does not receive the query due to a communications error, then no response is returned from the slave and the master device will eventually process a timeout condition for the query.
2. If the slave receives the query but detects a communication error (UART or CRC), then no response is returned from the slave and the master device will eventually process a timeout condition for the query.
3. If the slave receives the query without a communications error, and takes longer than the master's timeout setting, then no response is returned from the slave. The master device eventually processes a timeout condition for the query. To prevent this condition, the master timeout must be set longer than the maximum response time of the slave (10,000 milliseconds).
4. If the slave receives the query without a communications error but cannot process it due to reading or writing to a non-existent slave command register, then the slave returns an exception response message informing the master of the error.

The exception response message has two fields that differentiate it from a normal response. The first is the function code – byte 2. This code will have the high order bit set to a one (e.g., 0x83 for a read exception and 0x86 for a write exception). The second differentiating field is the exception code – byte 3. In addition, the total exception response length is 5 bytes rather than the normal message length.

**Table 11: Exception Response Packet**

Byte	Modbus Parameter	Range	Meaning
1	Slave Address	1 – 247	
2	Function Code	0x83 or 0x86	Read or Write
3	Exception Code	See Table Below	
4	CRC High	0x00 – 0xFF	
5	CRC Low	0x00 – 0xFF	

**Table 12: Exception Response Codes**

Code	Name	Description
1	Illegal Function Code	The function code received in the query is not an allowable action for the slave. This may be because the function code is only applicable to newer devices and was not implemented in the unit selected. It could also indicate that the slave is in the wrong state to process a request of this type, for example, it is not configured and is being asked to return register values.
2	Illegal Data Address	The data address received in the query is not an allowable address for the slave. More specifically, the combination of reference number and transfer length is invalid. For a controller with 100 registers, the PDU addresses the first register as 0, and the last one as 99. If a request is submitted with a starting register address of 96 and a quantity of registers of 4, then this request will successfully operate (address-wise at least) on registers 96, 97, 98, 99. If a request is submitted with a starting register address of 96 and a quantity of registers of 5, then this request will fail with Exception Code 0x02 “Illegal Data Address” since it attempts to operate on registers 96, 97, 98, 99 and 100, and there is no register with address 100.
3	Illegal Data Value	A value contained in the query data field is not an allowable value for slave. This indicates a fault in the structure of the remainder of a complex request, such as that the implied length is incorrect. It specifically does NOT mean that a data item submitted for storage in a register has a value outside the expectation of the application program, since the MODBUS protocol is unaware of the significance
4	Slave Device Failure	An unrecoverable error occurred while the slave was attempting to perform the requested action.

### 7.2.2 Modbus Command Register Definitions

The Command Register definitions for the HY-ALERTA 5000 Series Hydrogen Sensors are identified in Table 13.

**Note:** When reading registers containing 32 or 64-bit integers the user must read the high order word first, then the lower order word(s). Reading of the high order word causes the low order word to be saved in a temporary location for the next register read. The second register is then automatically read from the temporary location by the firmware. Likewise, with a write, the high value is stored until the second value is received, at which time, both values are written to the instrument.

**Table 13: Command Register Locations**

Register	Parameter	Function	Data Type	Data Range	Access
<b>Measurements</b>					
0	Hydrogen, ppm H <sub>2</sub> <small>Note: To convert ppm to %, divide ppm reading by 10,000</small>	High word	32-bit binary number	0 to 20,000,000	R
1		Low word			
2-6	Reserved for future use				
7	PCB Temperature, Celsius	x100 scale; 100 offset (T=V/100-100)	16-bit binary number	-100 to +200	R
8-30	Reserved for future use				

Register	Parameter	Function	Data Type	Data Range	Access
Information					
31-40	Model Number		ASCII String		R
41-50	Product Serial Number		ASCII String		R
51-60	Sensor Serial Number		ASCII String		R
61-70	Sensor Board Serial Number		ASCII String		R
71-80	Reserved for future use				
81	Manufacturing Date	high byte: Month low byte: Day	32-bit binary value		R
82		Year			
83	Factory Calibration Date	high byte: Month low byte: Day	32-bit binary value		R
84		Year			
85-86	Reserved for future use				
87	Field Calibration Date	high byte: Month low byte: Day	32-bit binary value		R
88		Year			
89-98	Firmware Revision		ASCII String		R
99-110	Reserved for future use				
Status/Error Information					
111	Status	Refer to section 8.2.3	16-bit binary flags	Table 14: Unit Status	R
112	Error Status	Refer to section 8.2.7.2 High word	32-bit binary flags	Table 15: Error Status	R
113		Low word			
114-120	Reserved for future use				

Register	Parameter	Function	Data Type	Data Range	Access
Calibration Functions					
121-125	Reserved for future use				
126	Calibration Gas 1, ppm H <sub>2</sub> <small>Note: To convert % to ppm, multiply % value by 10,000</small>	High word	32-bit binary number	0 to 1,000,000	R/W
127		Low word			
128	Calibration Complete/Date	High byte: Month Low byte: Day	32-bit binary value		R/W
129		Year			
130	Calibration Gas 2, ppm H <sub>2</sub> <small>Note: To convert % to ppm, multiply % value by 10,000</small>	High word	32-bit binary number	0 to 1,000,000	R/W
131		Low word			
132	Field Cal start command	Gas exposure duration (minutes)	16-bit binary number	60 to 1440	W
133	Field Cal abort command	Abort Field Cal or clear data	8-bit binary number	1: abort 2: clear	W
134	Field Cal get status command	High byte: Error code Low byte: Status	16-bit binary number	Table 5 and Table 6	R
135	Gas start command	Gas 1 or 2 started	8-bit binary number	1: gas 1 2: gas 2	W
Configuration Settings					
136-149	Reserved for future use				
150	Set Unit ID		8-bit binary number	1 to 247	R/W
151-158	Reserved for future use				
159	Stop Bit Selection		16-bit binary number	1 (stop bit = 1) 2 (stop bit = 2)	R/W
160	Baud Rate		8-bit binary number	1 = 9600 2 = 14400 3 = 19200 4 = 38400 5 = 57600 6 = 115200	R/W
161-174	Reserved for future use				

Register	Parameter	Function	Data Type	Data Range	Access	
Diagnostics						
175	Month / Year	Date & Time; read register 175 first; order high-byte / low- byte; add 2000 to year (64-bit)	Two Bytes		R/W	
176	Hour / Day		Two Bytes		R/W	
177	Second / Minute		Two Bytes		R/W	
178	Millisecond		16-bit binary number		R/W	
179-200	Reserved for future use					
User Information						
201-210	User ID #1	Must start reading from low address; Must write low and high addresses to save string	ASCII String		R/W	
211-220	User ID #2		ASCII String		R/W	
221-230	User ID #3		ASCII String		R/W	
231-255	Reserved for future use					

### 7.2.3 Hydrogen Measurement

The HY-ALERTA 5000 Series Hydrogen Sensors report the most recent hydrogen measurement in registers 0-1. The 32-bit unsigned integer value is not scaled and reports the integer value of hydrogen in ppm H<sub>2</sub>.

Note: Read the device status in register 111 bit 15 to determine if the device is ready. The hydrogen value is zero until the ready bit is set.

### 7.2.4 Temperature Measurement

The HY-ALERTA 5000 Series Hydrogen Sensors monitor the internal electronics temperature. The temperature is reported as a scaled 16-bit unsigned integer in degrees Celsius. Dividing the integer value by 100 and subtracting 100 will provide the measured temperature with 2 decimal places.

#### 7.2.4.1 PCB Temperature

PCB temperature is reported in register 7. This is the internal temperature of the electronics enclosure which must not exceed 105°C.

This is a good register to read during installation and communication testing because the value is always valid and frequently changes.

### 7.2.5 ASCII Strings

HY-ALERTA 5000 Series Hydrogen Sensor information is available as ASCII strings terminated with a zero byte (0x00). Each string can be up to 19 characters long with 2 characters per Modbus register. Use the read holding register function and read ten registers, each byte is an ASCII character.

#### 7.2.5.1 Model Number

The model number is in registers 31-40.

#### 7.2.5.2 Product Serial Number

The product serial number is in registers 41-50.

#### 7.2.5.3 Sensor Serial Number

The sensor serial number is in registers 51-60.

#### **7.2.5.4 Sensor Board Serial Number**

The sensor board serial Number is in registers 61-70.

#### **7.2.5.5 Firmware Revision**

The firmware revision is in registers 89-98, using format x:y:z; example 3:5:A

- x is the major revision
- y is the minor revision
- z is the product designator

#### **7.2.6 Date Register Format**

Registers that report a date value are encoded as follows.

- High word, high byte is the Month
- High word, low byte is the Day
- Low word is the Year

##### **7.2.6.1 Manufacturing Date**

Original manufacturing date is in registers 81,82

##### **7.2.6.2 Factory Calibration Date**

Last factory calibration date is in registers 83,84

#### **7.2.7 Status and Error Information**

The HY-ALERTA 5000 Series Hydrogen Sensors provide status and error information for the user to determine if it is operating normally.

##### **7.2.7.1 Unit Status**

Unit status information is maintained in Modbus register 111. The bit map for this status word is described below:

**Table 16: Unit Status**

<b>Bit #</b>	<b>Description</b>
15	Unit Ready, hydrogen readings are valid
14	New measurement data available, auto clear after register read
13	Unlisted bits are not used and may be 0 or 1.
12	Error, indicates an unrecoverable error occurred, read Reg 112,113 for more information
0-11	Unlisted bits are not used and may be 0 or 1.

### 7.2.7.2 Error Status

When the error flag (bit 12) of the Unit Status register 111 is set, the 32-bit register 112,13 has more information about what is causing the error. The bit map is shown below.

**Table 17: Error Status**

Bit #	Hex Value	Description
31	0x8000 0000	Sensor – Heater fault
30	0x4000 0000	Sensor – Temperature Sensor Fault
29	0x2000 0000	Sensor – Hydrogen Sensor Fault
3-28	0x1000 0000 - 0x0000 0008	Unlisted bits are not used and may be 0 or 1.
2	0x0000 0004	PCB Temperature greater than 105C
1	0x0000 0002	Required data not available
0	0x0000 0001	Configuration data not valid

### 7.2.8 Set Unit ID

The Modbus ID is reported or set in register 150. Reading this register is used to confirm that the selected ID is in use. Writing the desired ID to register 150 will set the unit to the specified ID. The device ID can range from 1 to 247 or as limited by the Modbus master. Note that if the current device ID is unknown then writing the desired ID to device 0 will broadcast the ID to all connected HY-ALERTA 5000 Series Hydrogen Sensor devices.

Preparing multiple units to share a common RS485 bus is accomplished by connecting one unit at a time to a Modbus controller and writing the desired ID for that unit to register 150 at device ID 0.

For PC-based configuration, use ComTest Pro from [www.BaseBlock.com](http://www.BaseBlock.com) for a Modbus controller. The device must be power cycled for the new ID to take effect. It is recommended that each device is labeled with the new device ID.

A simple procedure to configure multiple units is as follows:

- 1) Disconnect all units from the RS485 cable
- 2) Connect first unit to the RS485 cable
- 3) Use the Modbus Controller to write a single holding register (function 6) to register 150, device 0, with the desired ID for the connected unit
- 4) Wait up to 10 seconds for the Modbus response
- 5) Disconnect this unit and connect the next one to the RS485 cable
- 6) Repeat steps 3, 4, and 5 until all units are configured
- 7) Attach all units to the RS485 cable and read register 150 from each of the configured devices

### 7.2.9 Stop Bit Selection

To select which stop bit to use in the RS485 communication port settings, write a 1 or 2 to Modbus register **159** (default selection is **1**). **Note:** If changing to **2**, you **may** need to write a **3** to register **160** (baud rate – 19,200) as well and power-cycle the HY-ALERTA 5000 Series Hydrogen Sensor.

### 7.2.10 Baud Rate

The HY-ALERTA 5000 Series Hydrogen Sensor RS485 baud rate can be modified from the default 19,200 baud by writing to Modbus register 160 with the number corresponding to the desired value in the table below. The device must be power cycled for the new baud rate to take effect.

**Table 18: Baud Rate**

Number	Description
1	9,600 baud
2	14,400 baud
3	19,200 baud
4	38,400 baud
5	57,600 baud
6	115,200 baud

### 7.2.11 Real time Clock

HY-ALERTA 5000 Series Hydrogen Sensors have an internal real-time clock with backup power provided by a super capacitor. Depending on temperature, the backup power will last a few months in storage. During installation or after long periods of no power to the unit, the real-time clock should be set to the current date and time.

To set the date and time, write to registers 175, 176, 177, then 178; the time is saved when register 178 is written.

To get the date and time, read registers 175, 176, 177, then 178; the time is captured when register 175 is read.

#### 7.2.11.1 Month / Year

The month and year are in register 175; month in the high byte, year (from 2000) is in the low byte.

#### 7.2.11.2 Hour / Day

The hour and day are in register 176; hour (24-hour format) in the high byte, day in the low byte.

#### 7.2.11.3 Second / Minute

The second and Minute are in register 177; second in the high byte, minute in the low byte.

#### 7.2.11.4 Millisecond

Milliseconds are in register 178.

### 7.2.12 User Information

The HY-ALERTA 5000 Series Hydrogen Sensors provide three ASCII strings that the user can program to indicate the where the sensor is installed. Each string can be up to 20 characters including null termination.

#### 7.2.12.1 User ID #1

The user string #1 is saved in registers 201 through 210.

#### 7.2.12.2 User ID #2

The user string #2 is saved in registers 211 through 220.

#### 7.2.12.3 User ID #3

The user string #3 is saved in registers 231 through 230.

## 8 Firmware Upgrade

The HY-ALERTA 5000 Series Hydrogen Sensor firmware is field-upgradable. Instructions and PC software will be provided by H2scan as needed.


Copy the firmware upgrade files to a directory on the PC. (**NOTE:** The latest available firmware will be made available)

- **g5RX\_YA.exe** – sensor firmware binary file
- **g5RX\_YA.bat** – batch file


**Example:** g5R3\_5A.exe and g5R3\_5A.bat

7-Step procedure to upgrade firmware:

- 1) Copy the files below to your local directory.




- 2) Connect the HY-ALERTA 5000 Series Hydrogen Sensor to your PC using a USB to RS485 adapter.
- 3) Right-click on g5RX\_YA.bat and then select Edit.




- 4) You will see the port information below. Update the com port, the device address and the baud rate you are using to connect to the HY-ALERTA™ 5000 Series Hydrogen Sensor. Save the file.

ComPort = comX → change to the com port being used  
 Device Address = 1 → default address is 1  
 Device Baud Rate = 19200 → default baud rate is 19200



- 5) Double-click on g5RX\_YA.bat. You will see the following information as the firmware is being upgraded.



```
C:\Windows\system32\cmd.exe
C:\Users\jgoodheart\Desktop\Firmware\Gen5\3.5.A (hex -333A 353A 4100)>g5R3_5A.exe com4 1 19200

Gen5 SFU Utility Software Version 2.00
\\.\com4, baudrate = 19200, devID = 1, size=3ffc8, crc=f7a3
1. Open Serial Comm Port.
nothing read
2. Check if Gen5 is in Modbus or CLI mode.


*Gen5 is in Modbus mode.
3. Switching to CLI mode.

*FW=3:5:A, PORRST>
got prompt!
FW=3:5:A,4. Ready for code download. Begin.

*
*
*aaa 3ffc8 f7a3
*abb 8000 df7e
....
*abb 8000 df7e

*abb 8000 23c2
```

6) The upgrade process is complete when you see the window below.



```
C:\Windows\system32\cmd.exe
....
*abb 8000 e6f8
*abb 8000 8151
....
*abb 8000 8151
*abb 8000 85fd
....
*abb 8000 85fd
*abb 7fc8 e0aa
...
*abb 7fc8 e0aa
6. Code download is done
*acc 3ffc8 f7a3

SUCCESS: final size = 0x3ffc8, crc = 0xf7a3
7. Close Serial Comm Port.

C:\Users\jgoodheart\Desktop\Firmware\Gen5\3.5.A (hex -333A 353A 4100)>pause
Press any key to continue . . .
```

7) To complete the process, please power cycle the HY-ALERTA 5000 Series Hydrogen Sensor.

## 9 Principles of Operation

### 9.1 Effect of Pressure

The HY-ALERTA 5000 Series Hydrogen Sensors are hydrogen-specific and measure the partial pressure of only hydrogen in a gas stream. Changes in gas pressure will affect the hydrogen partial pressure and therefore the measured output as well.

Example:

At 1.0 atmosphere absolute, a 2% H<sub>2</sub> by volume mixture will be reported as 2% H<sub>2</sub> by the analyzer. At 0.9 atmospheres absolute the reading will decrease to 1.8% H<sub>2</sub> and 2.0 atmospheres absolute will result in a reading of 4% H<sub>2</sub>.

The analyzers are factory calibrated at 1.0 atmosphere absolute. For operation at other pressures, a 2-point field calibration must be performed upon installation to ensure the H<sub>2</sub> reading provides the correct volumetric concentration of H<sub>2</sub>. (Refer to section 7 for detailed instructions.) Alternatively, if the pressure is known or actively measured, the volumetric H<sub>2</sub> concentration can be calculated using the equation:

$$H_{2vol} = \frac{H_{2read}}{P_{ata}}$$

Where  $P_{ata}$  is the known or measured pressure in atmospheres absolute and  $H_{2read}$  is the H<sub>2</sub> reading from the analyzer. It is important to note that any error in the pressure measurement will translate as error in the H<sub>2</sub> concentration measurement. For best performance, a nonvarying pressure at 1.0 atmosphere absolute is recommended.

## 9.2 Pressure Normalized Specifications

Because the sensing element is sensitive to the partial pressure of hydrogen rather than the volumetric or molecular concentration, the operating pressure of the analyzer can significantly affect its performance. The reported hydrogen is given as a volumetric concentration and the pressure is assumed to be constant. The specifications listed in Section 3 Specifications are given as volumetric concentrations at 1 atmosphere absolute.

As described in Section 10.1 there is a simple physical relationship between volumetric concentration and partial pressure. A gas containing 10% hydrogen by volume at 1 atmosphere absolute can be thought of instead as having a partial pressure of 0.1 atmospheres absolute of hydrogen, and 10% hydrogen by volume at 2 atmospheres absolute would correspond to 0.2 atmospheres absolute of hydrogen. (For more information on the subject refer to Dalton's Law.)

The table below describes the specifications for each product in units of atmospheres absolute. These specifications assume a dry measurand, an ambient temperature of 25°C, and are in addition to any errors in the calibration gases used.

**Table 19: Partial Pressure Limits**

MODEL	H <sub>2</sub> Range		Lower Detection Limit (LDL)
	Low	High	
HY-ALERTA 5020	0	0.05	0.004
HY-ALERTA 5021	0	0.05	0.004

Units: atmospheres absolute (ata)

**Example:** A gas containing up to 1% H<sub>2</sub> by volume can be monitored by a HY-ALERTA 5020 at up to 5 atmospheres absolute. At this operating pressure, the analyzer would see up to 0.05 atmospheres absolute of H<sub>2</sub> which is within the limits listed in the table above. However, if the operating pressure were to be raised to 6 atmospheres absolute, the 1% H<sub>2</sub> gas would exceed the limit at 0.06 atmospheres absolute of hydrogen.

## 9.3 Sensor Behavior

An internal algorithm is utilized to maintain long term performance of the sensors and remove the need for periodic field calibrations. For this algorithm to run properly the analyzers must be powered on continuously. If the analyzer is powered only for short durations (e.g. <2 hours), the algorithm will not function properly, and the accuracy of the sensor may not meet the product specifications.

The sensors in the HY-ALERTA models 5020 and 5021 are designed to monitor the H<sub>2</sub> concentration in air of an area, building, battery room, cabinet, etc. These sensors can operate in an environment where H<sub>2</sub> is normally not present for extended periods.

It is important to note that the sensors should be *powered on* in a hydrogen-free environment containing oxygen (such as air) for most of their operation. If the sensors are only powered on during periods of hydrogen

measurement their performance will be degraded. To restore performance, power the sensor in a hydrogen-free environment containing oxygen for at least 24 hours then perform a 2-point field calibration.

#### **9.4 Non-reducing and Non-oxidizing Atmospheres**

If the sensor is left powered on in a gas that contains no hydrogen and no oxygen for periods of days or longer it will experience drift. An internal algorithm will ensure the reported H<sub>2</sub> remains at 0. However, when the sensor is eventually exposed to hydrogen (above the lower detection limit), the initial response may be inaccurate. An internal algorithm will restore the accuracy if exposed to any constant hydrogen concentration (within the range of the model) for at least 4 hours.

#### **9.5 Diurnal Effect**

Temperature changes throughout the day may induce small changes to the H<sub>2</sub> reading of the analyzer. As explained in section 10.1 Effect of Pressure, the sensor measures the number of hydrogen molecules present which means pressure changes (brought on by temperature fluctuations) can create minor fluctuations in the reported H<sub>2</sub> concentration. While the software in the analyzer corrects mostly for these effects it may still be observed in final reading.

#### **9.6 Response Time**

The response of the sensor to a change in hydrogen follows a typical “s-curve”, 1<sup>st</sup> order system response in most cases. As a change in the partial pressure of hydrogen around the sensor die interacts with the palladium lattice on the sensing element, its electrical properties change immediately. If the sensor is not oxidized, this change will be reflected in the analyzer’s hydrogen output within a few seconds.

For applications where the sensor is exposed to a gas containing oxygen and no hydrogen (or less than ~500ppm of hydrogen), an oxide layer will form on the surface of the sensor. This oxide layer will be quickly removed in the presence of hydrogen, but it must be fully removed before the sensor will begin showing a response to the hydrogen. The duration of this process depends on how long the sensor was exposed to oxygen before the initial hydrogen exposure as well as the hydrogen concentration.

In general, sensors are considered fully oxidized after being powered on in a gas containing oxygen (and no hydrogen) for more than a week. As mentioned previously, the time it takes to remove this oxide layer in order to respond to the hydrogen depends on the hydrogen concentration. Assuming an operating pressure around 1 atmosphere absolute, a fully oxidized sensor being exposed to 2% hydrogen should have its oxide layer removed (and thus begin showing a change in the reported H<sub>2</sub>) within 30 seconds and be 90% through the transition within 60 seconds. If, however, the fully oxidized sensor was exposed to 0.4% hydrogen, the oxide layer may take over 60 seconds to be fully removed and start reporting an increase in H<sub>2</sub>. A response to 4% H<sub>2</sub> (the LEL) will remove the oxide layer in just a few seconds.

If the gas contains oxygen but small amounts of hydrogen such as 500ppm, this may be enough to prevent the formation of an oxide layer and will allow the sensor to respond more quickly to an increase in hydrogen.

When a fast response time is critical or when measuring response time, it is important to understand that the sensor can only respond to the gas at the sensor. A transition from a first condition to a second may be instantaneous at the source, but there is always some transport time before the gas reaches with the sensor.

#### **9.7 Sensor Performance Evaluation (Response Time)**

When evaluating the response time of HY-ALERTA 5000 Series Hydrogen Sensors, the analyzers must be installed into a gas system such that the gas mixture can delivered to the sensing elements of the device and changed quickly. H2scan’s response time specifications are based on a step response transition from a first condition to a second condition **at the sensing elements**, where a condition consists of a gas with some hydrogen concentration diluted in some background gas at a temperature of 25°C and pressure of 1 atmosphere absolute. For the HY-ALERTA 5020 and 5021, the gas is transitioned from 0% H<sub>2</sub> (100% Air) to 3% H<sub>2</sub> in Air. The response time specification defines the maximum amount of time before the unit will report a hydrogen concentration of at least 2%.

H2scan uses a response time apparatus at the factory that is designed to instantaneously change the gas concentration from the first value to the second at the sensing elements. This is done by allowing the sensing elements to equilibrate in a very small volume of the first gas, applied with a low flow rate, then initiating the transition by stopping the flow of the first gas and opening the small volume to a significantly larger volume that contains the second gas. This larger volume is also maintained with a flow of the second gas. Because the volume

of the second gas is significantly greater than that of the first gas, as the two gases mix the concentration is effectively equivalent to the value of the second gas.

However, it may not be practical to construct an apparatus like this for an evaluation, so a sensor flow system may be required. When considering the mixture of gas in a system during a transition, a general rule is that each standard volume turnover will achieve 90% of the transition. For example, if the total system volume (including plumbing) is 1 standard liter and the flow rate is 1 slpm, then 1 minute after the start of the transition the gas would be estimated to comprise 10% the first condition and 90% the second. After 2 minutes this becomes 1% and 99%, and so on and so forth. The hydrogen measurement of the sensor is a function of the gas mixture *at the sensor*, so minimizing the response time and more accurately determining the response time of the sensor can be achieved by:

- Minimizing the standard volume of gas between the source and the sensor by:
  - Using a small diameter pipe
  - Minimizing the length of pipe
  - Reducing the gas pressure to (or near to) the desired operating pressure early in the flow path

**Note:** The recommended operating pressure is 1 atmosphere absolute. Deviating from this operating pressure will affect the hydrogen measurements of the analyzer. See section 10.1 Effect of Pressure.

- Increasing the volumetric gas flow rate

**Note:** A high flow rate in a system with small diameter pipe, small diameter pipe bends, changes in orifice size, etc., can affect the pressure or cause turbulent flow.

It is up to the end user to understand the principles of sample stream conditioning and to determine the ideal plumbing for the application. A constant, stable pressure and laminar, steady flow at the sensor is ideal.

## 9.8 Sensor Performance Evaluation (Accuracy)

When evaluating the accuracy of the HY-ALERTA 5021, it is important to note that the intended application for this product is battery room monitoring. In this application there may be a gradual increase in hydrogen in air over the course of hours, days, weeks, or even months, depending on the situation.

When the sensor is powered on in air that contains no hydrogen, the autocalibration algorithm will ensure the hydrogen reading remains at zero despite the development of an oxide layer on the surface of the sensing elements as described in section 10.3 Sensor Behavior. When the air contains small amounts of hydrogen, roughly between 500ppm and 1000ppm, that oxide layer will begin to deplete. At these low hydrogen levels this process may take several hours, but for higher concentrations, such as 5% H<sub>2</sub>, this could take just a few seconds. Once the oxide layer is depleted and the hydrogen levels increase to above the LDL (4000ppm H<sub>2</sub>), the autocalibration algorithm will work to ensure the measurements remain highly accurate for as long as the hydrogen concentration remains above the LDL. This algorithm is specially tuned to track gradual increases and decreases in hydrogen concentration in air over long periods.

When evaluating the accuracy of the HY-ALERTA 5021 Hydrogen Sensor, it is recommended that the sensing elements first be exposed to a gas containing at least 4000ppm H<sub>2</sub> for at least 2 hours to deplete the oxide layer. This will ensure that the sensor is tested similar to the expected operating mode in the field, where background levels of hydrogen will likely be present long before crossing the 4000ppm LDL threshold.

## 9.9 General Troubleshooting

Here are some general questions to answer if problems are encountered with the analyzer.

1. Is the analyzer reporting %H<sub>2</sub>?
  - Check all connections and flow paths.
  - Reduce process or calibration gas stream pressure to 1.0 atm and power cycle the analyzer.
2. Are the pressure and the flow stable?

- Unstable pressure or flow can cause the analyzer to behave erratically. Ensure the process gas stream is properly regulated. The analyzer is less sensitive to changes in flow but a flow rate of 1.0 slpm is recommended.

3. Is the sensor output stable?

- Expose the analyzer to a calibration gas (40-100% H<sub>2</sub> but not exceeding 10% H<sub>2</sub> in the model 5031) overnight while recording the H<sub>2</sub> reading.
- Plot the H<sub>2</sub> reading. (You may use spreadsheet software such as Microsoft Excel). Assuming a constant pressure, temperature, and flow rate, the H<sub>2</sub> line for the most recent data should be flat with no perturbations outside the defined specifications for the analyzer. If it is not yet stable, leave the analyzer exposed to the calibration gas for another 24 hours or until stability is observed.
- Once the sensor is stable, perform a 2-point field calibration (see section 7).
- If the sensor does not stabilize, contact H2scan for support.

4. Has the sensor been exposed to substances that may have damaged it?

- Exposure to liquids, acids, bases, or levels of H<sub>2</sub>S or CO that exceed the product limits will damage the sensor die.